

Determination of PCB in IRMM reference materials BCR-682 (mussel tissue) and BCR-718 (herring)

M. Hoek- van Nieuwenhuizen

Report number C038/11

IMARES Wageningen UR

(IMARES - Institute for Marine Resources & Ecosystem Studies)

Client: European Commission-IRMM
Barbara Raffaelli
Retieseweg 111
2440 Geel
Belgium

Publication date: 24th March

IMARES is:

- an independent, objective and authoritative institute that provides knowledge necessary for an integrated sustainable protection, exploitation and spatial use of the sea and coastal zones;
- an institute that provides knowledge necessary for an integrated sustainable protection, exploitation and spatial use of the sea and coastal zones;
- a key, proactive player in national and international marine networks (including ICES and EFARO).

P.O. Box 68
1970 AB IJmuiden

Phone: +31 (0)317 48 09 00

Fax: +31 (0)317 48 73 26

E-Mail: imares@wur.nl
www.imares.wur.nl

P.O. Box 77
4400 AB Yerseke

Phone: +31 (0)317 48 09 00

Fax: +31 (0)317 48 73 59

E-Mail: imares@wur.nl
www.imares.wur.nl

P.O. Box 57
1780 AB Den Helder

Phone: +31 (0)317 48 09 00

Fax: +31 (0)223 63 06 87

E-Mail: imares@wur.nl
www.imares.wur.nl

P.O. Box 167
1790 AD Den Burg
Texel
Phone: +31 (0)317 48 09 00
Fax: +31 (0)317 48 73 62
E-Mail:
imares@wur.nl
www.imares.wur.nl

© 2010 IMARES Wageningen UR

IMARES, institute of Stichting DLO
is registered in the Dutch trade
record nr. 09098104,
BTW nr. NL 806511618

The Management of IMARES is not responsible for resulting damage, as well as for damage resulting from the application of results or research obtained by IMARES, its clients or any claims related to the application of information found within its research. This report has been made on the request of the client and is wholly the client's property. This report may not be reproduced and/or published partially or in its entirety without the express written consent of the client.

A_4_3_2-V11.2

Contents

Contents	3
1. Introduction.....	4
2. Materials and Method	4
3. Results.....	4
4. Quality Assurance	5
Justification	7
Appendix A. Results of PCB in reference materials BCR-682 and BCR-718 in µg/kg fresh material	8

1. Introduction

The institute of Reference Materials and Measurements (IRMM) asked IMARES to analyse PCB in two different reference materials, LOT 1: BCR-682 (mussel tissue) and LOT 2: BCR-718 (herring).

2. Materials and Method

On the 15th of February 2011 IMARES received two different reference materials, four units each, from IRMM for the analyses of the following PCB:

LOT 1: BCR-682 (mussel tissue; units 15, 70, 115 and 339): CB28, 52, 118, 138+163, 149, 153, 170, 180.

LOT 2: BCR-718 (herring; units 3, 63, 150, 436): CB28, 52, 101, 105, 118, 128, 138+163, 149, 153, 156, 170, 180.

There was agreed upon that analyses should be performed in March 2011 (see e-mail d.d. 14th of January).

In each of the four units of a reference material three measurements were performed under repeatability conditions (12 determinations of each reference material), so 24 measurements of two different reference materials in one analytical run.

The PCB analyses were performed according to the following method (ISW 2.10.3.001):

Sample mixed with sodium sulphate (for removing moisture) is Soxhlet extracted for 6.5 hours with pentane/dichloromethane (1:1). After addition of 5 ml of iso-octane as a keeper, the extract is concentrated to about 5 ml in a rotary evaporator. Further clean-up and fractionation is then performed by putting the extract on top of a 15 g alumina · 9% H₂O column and elution with 170 ml n-pentane. After the addition of 2 ml of iso-octane as a keeper, the eluate is concentrated to about 2 ml in a rotary evaporator. The concentrate is transferred to a silica column (1.6 g silica · 2% H₂O) and eluted with 11 ml of iso-octane. Test determination by GC-ECD is then performed to estimate the concentration of the analytes in the eluate. If values are out of the range of the calibration curve, dilution or concentration of the eluate is performed. Final determination is then performed by gas chromatography (GC) coupled to electron capture detector (ECD) using two column system: Cp-Sil 8 (50 m long with internal diameter of 0.15 mm and film thickness of 0.20 µm) and Cp-Sil 19 (50 m long with internal diameter of 0.15 mm and film thickness of 0.20 µm).

All analyses of the requested PCB's are under accreditation by the Council for Accreditation in the matrix fish: CB28, 52, 101, 105, 118, 128, 138+163, 149, 153, 156, 170, 180 (Test laboratory code L097, Activity number 9). PCB138 will be reported as the sum of PCB138+163.

IMARES is registered as a Reference Lab at the Europese Commissie-Institute for Reference Materials and Measurements (IRMM) for the determination of PCB.

3. Results

The results given in this report apply only to the samples analyzed.

The results are given in appendix A. Concentrations are reported in µg/kg fresh material.

Decimal characters: Data is in derogation Dutch SI reported a decimal point (.) Instead of a comma (,).

The levels of the concentrations of the PCB in both reference materials BCR-682 and BCR-718 are very low in comparison of the levels of the PCB in our IRM cod liver 406 (see table 2).

4. Quality Assurance

IMARES utilises an ISO 9001:2008 certified quality management system (certificate number: 57846-2009-AQ-NLD-RvA). This certificate is valid until 15 December 2012. The organisation has been certified since 27 February 2001. The certification was issued by DNV Certification B.V. Furthermore, the chemical laboratory of the Environmental Division has NEN-AND-ISO/IEC 17025:2005 accreditation for test laboratories with number L097. This accreditation is valid until 27 March 2013 and was first issued on 27 March 1997. Accreditation was granted by the Council for Accreditation.

For every sequence of samples our results are tested on internal or certified reference materials (IRM/CRM). The results of the reference materials are collected in Quality Assurance Charts conform NPR 6603.

IMARES has a prominent role in Quasimeme preparing reference materials for certifying purposes and organizing international interlaboratory tests for contaminants in environmental samples. The accuracy of our methods is tested by attendance to Performance Laboratory Studies (Quasimeme).

Our **Quasimeme results** for the requested PCB are given in table 1.

Table 1. Quasimeme results over several years

Component	Z-score			
	satisfied	unsatisfied	questionable	total
CB28	38	2	2	42
CB52	48	2	4	54
CB101	43	7	5	55
CB105	37	2	11	50
CB118	45	4	5	54
CB128	<i>not specified</i>			
CB138+163	42	2	2	46
CB149	<i>not specified</i>			
CB153	51	2	1	54
CB156	34	3	0	37
CB170	<i>not specified</i>			
CB180	46	3	2	51

The **repeatability** and **trueness** of the requested PCB in an internal reference material (IRM) cod liver 406 are given in table 2.

Table 2. Repeatability and trueness

Component	repeatability CV(%) n=9	trueness*	
		J(%)	n
CB28	2.8 at a level of 45.9 µg/kg	99.9	8
CB52	2.1 at a level of 160 µg/kg	91.0	8
CB101	2.1 at a level of 410 µg/kg	109.8	8
CB105	5.8 at a level of 138 µg/kg	90.4	8
CB118	4.0 at a level of 461 µg/kg	98.4	10
CB128	2.3 at a level of 88 µg/kg	<i>not specified</i> <i>because of sum of CB138+163</i> <i>no trueness specified</i>	
CB138+163	3.1 at a level of 808 µg/kg	89.4	6
CB149	2.5 at a level of 220 µg/kg	97.6	10
CB153	2.0 at a level of 1128 µg/kg	100.6	9
CB156	2.7 at a level of 60 µg/kg	<i>not specified</i>	
CB170	2.9 at a level of 140 µg/kg	<i>not specified</i>	
CB180	3.6 at a level of 309 µg/kg	96.4	8

The **standard uncertainty** and the **expanded uncertainty** according to NEN 7779 are mentioned in table 3.

Table 3. Standard and expanded uncertainty

Component	standard uncertainty (%) (%)	expanded uncertainty (%)	n	Accreditation
PCB28	20.5	40.9	36	Q
PCB52	17.3	34.6	45	Q
PCB101	16.3	32.6	42	Q
PCB105	24.6	49.2	43	Q
PCB118	16.8	33.6	44	Q
PCB128	not defined	not defined		Q
PCB138+163	16.7	33.4	45	Q
PCB149	not defined	not defined		Q
PCB153	9.9	19.8	45	Q
PCB156	19.1	38.2	30	Q
PCB170	not defined	not defined		Q
PCB180	18.1	36.5	43	Q

Justification

Rapport C038/11

Project Number: **4305107601**

The scientific quality of this report has been peer reviewed by the a colleague scientist and the head of the department of IMARES.

Approved: Dr. Ir. M.J.J. Kotterman

Project leader

Signature:



Date:

24th March

Approved: Drs. J.H.M. Schobben

Head of the department Environment



Signature:

24th March

**Appendix A. Results of PCB in reference materials BCR-682 and BCR-718 in
µg/kg fresh material**

Appendix A. Results of PCB in reference materials BCR-682 and BCR-718 in µg/kg fresh material

LIMS number	IRMM Unit ID-number	replicate	reference material	date of analysis	CB28 µg/kg	CB52 µg/kg	CB101 µg/kg	CB105 µg/kg	CB118 µg/kg	CB128 µg/kg	CB138+163 µg/kg	CB149 µg/kg	CB153 µg/kg	CB156 µg/kg	CB170 µg/kg	CB180 µg/kg
2011/0693	0682/15	1	BCR-682	2th of March 2011	0.36	0.86	n.a.	2.27	n.a.	6.00	5.05	8.86	n.a.	0.21	0.96	
2011/0694		2	BCR-682	2th of March 2011	0.35	0.81	n.a.	n.a.	2.27	n.a.	5.83	5.06	8.88	n.a.	0.21	0.92
2011/0695		3	BCR-682	2th of March 2011	0.40	0.80	n.a.	n.a.	2.29	n.a.	5.77	4.98	8.83	n.a.	0.21	0.97
2011/0696	0682/70	1	BCR-682	2th of March 2011	0.38	0.92	n.a.	n.a.	2.49	n.a.	6.45	5.43	9.79	n.a.	0.21	0.99
2011/0697		2	BCR-682	2th of March 2011	0.36	0.87	n.a.	n.a.	2.41	n.a.	6.33	5.26	9.51	n.a.	0.20	0.95
2011/0698		3	BCR-682	2th of March 2011	0.38	0.89	n.a.	n.a.	2.47	n.a.	6.58	5.24	9.46	n.a.	0.21	0.98
2011/0699	0682/115	1	BCR-682	2th of March 2011	0.33	0.77	n.a.	n.a.	2.28	n.a.	5.80	4.75	8.60	n.a.	0.21	0.91
2011/0700		2	BCR-682	2th of March 2011	0.35	0.82	n.a.	n.a.	2.23	n.a.	5.87	4.86	8.88	n.a.	0.22	0.90
2011/0701		3	BCR-682	2th of March 2011	0.39	0.83	n.a.	n.a.	2.32	n.a.	6.05	4.85	8.64	n.a.	0.22	0.91
2011/0702	0682/339	1	BCR-682	2th of March 2011	0.38	0.93	n.a.	n.a.	2.27	n.a.	6.16	5.20	9.24	n.a.	0.20	0.93
2011/0703		2	BCR-682	2th of March 2011	0.37	0.90	n.a.	n.a.	2.43	n.a.	6.18	5.22	9.39	n.a.	0.20	0.96
2011/0704		3	BCR-682	2th of March 2011	0.39	0.98	n.a.	n.a.	2.41	n.a.	6.25	5.23	9.19	n.a.	0.20	0.89
2011/0705	0718/3	1	BCR-718	2th of March 2011	0.50	1.21	2.08	0.67	2.06	0.68	3.21	2.61	4.41	0.20	0.33	0.86
2011/0706		2	BCR-718	2th of March 2011	0.46	1.07	2.01	0.64	1.97	0.64	2.95	2.52	4.38	0.21	0.33	0.85
2011/0707		3	BCR-718	2th of March 2011	0.47	1.16	2.07	0.66	2.05	0.68	3.28	2.53	4.41	0.22	0.28	0.86
2011/0708	0718/63	1	BCR-718	2th of March 2011	0.62	1.27	2.07	0.59	1.91	0.60	3.04	2.63	4.18	0.22	0.30	0.78
2011/0709		2	BCR-718	2th of March 2011	0.43	1.01	1.78	0.61	1.70	0.58	2.80	2.42	3.88	0.20	0.29	0.76
2011/0710		3	BCR-718	2th of March 2011	0.45	1.04	1.84	0.60	1.77	0.58	2.91	2.44	3.94	0.21	0.29	0.76
2011/0711	0718/150	1	BCR-718	2th of March 2011	0.42	1.00	1.72	0.57	1.67	0.55	2.64	2.34	3.80	0.18	0.27	0.69
2011/0712		2	BCR-718	2th of March 2011	0.47	1.12	2.05	0.64	1.90	0.67	3.16	2.66	4.28	0.22	0.31	0.81
2011/0713		3	BCR-718	2th of March 2011	0.43	1.02	1.90	0.63	1.91	0.61	2.91	2.51	4.26	0.21	0.31	0.81
2011/0714	0718/436	1	BCR-718	2th of March 2011	0.43	1.08	1.92	0.66	1.91	0.61	3.00	2.54	4.25	0.23	0.31	0.79
2011/0715		2	BCR-718	2th of March 2011	0.50	1.06	1.89	0.64	1.91	0.63	3.09	2.57	4.37	0.24	0.31	0.85
2011/0716		3	BCR-718	2th of March 2011	0.46	1.02	1.89	0.63	1.89	0.61	3.07	2.50	4.14	0.22	0.31	0.82

n.a. = not to be analyzed

Remark: the concentrations are reported with 2 digits on request of the client

Appendix A. Results of PCB in reference materials BCR-682 and BCR-718 in µg/kg fresh material

LIMS number	IRMM Unit ID-number	replicate	reference material	date of analysis	CB28 µg/kg	CB52 µg/kg	CB101 µg/kg	CB105 µg/kg	CB118 µg/kg	CB128 µg/kg	CB138+163 µg/kg	CB149 µg/kg	CB153 µg/kg	CB156 µg/kg	CB170 µg/kg	CB180 µg/kg
2011/0693	0682/15	1	BCR-682	2th of March 2011	0.36	0.86	n.a.	n.a.	2.27	n.a.	6.00	5.05	8.86	n.a	0.21	0.96
2011/0694		2	BCR-682	2th of March 2011	0.35	0.81	n.a.	n.a.	2.27	n.a.	5.83	5.06	8.88	n.a	0.21	0.92
2011/0695		3	BCR-682	2th of March 2011	0.40	0.80	n.a.	n.a.	2.29	n.a.	5.77	4.98	8.83	n.a	0.21	0.97
2011/0696	0682/70	1	BCR-682	2th of March 2011	0.38	0.92	n.a.	n.a.	2.49	n.a.	6.45	5.43	9.79	n.a	0.21	0.99
2011/0697		2	BCR-682	2th of March 2011	0.36	0.87	n.a.	n.a.	2.41	n.a.	6.33	5.26	9.51	n.a	0.20	0.95
2011/0698		3	BCR-682	2th of March 2011	0.38	0.89	n.a.	n.a.	2.47	n.a.	6.58	5.24	9.46	n.a	0.21	0.98
2011/0699	0682/115	1	BCR-682	2th of March 2011	0.33	0.77	n.a.	n.a.	2.28	n.a.	5.80	4.75	8.60	n.a	0.21	0.91
2011/0700		2	BCR-682	2th of March 2011	0.35	0.82	n.a.	n.a.	2.23	n.a.	5.87	4.86	8.88	n.a	0.22	0.90
2011/0701		3	BCR-682	2th of March 2011	0.39	0.83	n.a.	n.a.	2.32	n.a.	6.05	4.85	8.64	n.a	0.22	0.91
2011/0702	0682/339	1	BCR-682	2th of March 2011	0.38	0.93	n.a.	n.a.	2.27	n.a.	6.16	5.20	9.24	n.a	0.20	0.93
2011/0703		2	BCR-682	2th of March 2011	0.37	0.90	n.a.	n.a.	2.43	n.a.	6.18	5.22	9.39	n.a	0.20	0.96
2011/0704		3	BCR-682	2th of March 2011	0.39	0.98	n.a.	n.a.	2.41	n.a.	6.25	5.23	9.19	n.a	0.20	0.89
2011/0705	0718/3	1	BCR-718	2th of March 2011	0.50	1.21	2.08	0.67	2.06	0.68	3.21	2.61	4.41	0.20	0.33	0.86
2011/0706		2	BCR-718	2th of March 2011	0.46	1.07	2.01	0.64	1.97	0.64	2.95	2.52	4.38	0.21	0.33	0.85
2011/0707		3	BCR-718	2th of March 2011	0.47	1.16	2.07	0.66	2.05	0.68	3.28	2.53	4.41	0.22	0.28	0.86
2011/0708	0718/63	1	BCR-718	2th of March 2011	0.62	1.27	2.07	0.59	1.91	0.60	3.04	2.63	4.18	0.22	0.30	0.78
2011/0709		2	BCR-718	2th of March 2011	0.43	1.01	1.78	0.61	1.70	0.58	2.80	2.42	3.88	0.20	0.29	0.76
2011/0710		3	BCR-718	2th of March 2011	0.45	1.04	1.84	0.60	1.77	0.58	2.91	2.44	3.94	0.21	0.29	0.76
2011/0711	0718/150	1	BCR-718	2th of March 2011	0.42	1.00	1.72	0.57	1.67	0.55	2.64	2.34	3.80	0.18	0.27	0.69
2011/0712		2	BCR-718	2th of March 2011	0.47	1.12	2.05	0.64	1.90	0.67	3.16	2.66	4.28	0.22	0.31	0.81
2011/0713		3	BCR-718	2th of March 2011	0.43	1.02	1.90	0.63	1.91	0.61	2.91	2.51	4.26	0.21	0.31	0.81
2011/0714	0718/436	1	BCR-718	2th of March 2011	0.43	1.08	1.92	0.66	1.91	0.61	3.00	2.54	4.25	0.23	0.31	0.79
2011/0715		2	BCR-718	2th of March 2011	0.50	1.06	1.89	0.64	1.91	0.63	3.09	2.57	4.37	0.24	0.31	0.85
2011/0716		3	BCR-718	2th of March 2011	0.46	1.02	1.89	0.63	1.89	0.61	3.07	2.50	4.14	0.22	0.31	0.82

n.a. = not to be analyzed

Remark: the concentrations are reported with 2 digits on request of the client